



3EC International a.s., Hraničná 18, 821 05 Bratislava, Slovakia
Notified Body No. 2265

EC CERTIFICATE

No. 2018-MDD/QS-012

issued in compliance with the Council Directive 93/42/EEC as amended by 2007/47/EC, which is implemented by the Slovak Government Decree No. 582/2008 Coll. as amended by 215/2013 Coll. certifies that the medical device of Class IIb,

Infant Warmer
Brand name: Wallaby Warmer
Model: V1

manufactured by company

Medical Technology Transfer and Services Company Limited
House No. 26, Alley 41, An Duong Vuong Street, Phu Thuong Ward, Tay Ho District,
Hanoi City, Vietnam


is manufactured under conditions fulfilling the quality system requirements of Annex II, excluding (4), of the Directive 93/42/EEC as amended by 2007/47/EC.

The Notified Body No. 2265 has performed an audit of the above device quality system. The full quality assurance system has been assessed and found that it meets the requirements above. The quality system is subject to continuous surveillance according to Annex II, Sections 3.3, and 5, of the Directive 93/42/EEC as amended by 2007/47/EC. The detailed description of the system, requirements and measures applied by the manufacturer are presented in the Audit Report No. 310338, and the Final protocol No. 310338/2018 that is enclosed to this certificate.

This certificate is issued under the following conditions:

It applies only to the quality system maintained in the manufacture of the above referenced model of medical device and it does not substitute the design or type-examination procedures, if requested. The certificate remains valid until the manufacturing conditions or the quality system are changed but until December 14th, 2019 at the latest. The certificate validity is conditional upon positive results of regular surveillance audits and fulfilment of relevant legal and other requirements by manufacturer.




Dr. Katarína Srdošová
Responsible to act on behalf of NB 2265

In Bratislava, on September 8th, 2018